

德国 HYDRO-BIOS 公司--LIMNOS 水样采集器

LIMNOS Water Sampler



LIMNOS 水样采集器(玻璃采水器)是一款结构简单、操作便捷的水样采集设备,适用于 30m 以内的水深采样。LIMNOS 水样采集器由 PVC 固定器和可更换的 1000ml 玻璃采样瓶组成。

为了防止样品被表面水污染,将处在关闭状态的 LIMNOS 水样采集器系在一根线缆上,沉入水中。当达到预期深度时,使锤将会撞击铁砧,这时硅树脂管将会弹起,允许水样进入样品瓶,同时将空气排出。

LIMNOS 水样采集器(玻璃采水器)订购信息:

436 140 LIMNOS 水样采集器

适用于 30m 水深,每套带 2 个 1L 玻璃采样瓶和一个塑料包被的使锤

长: 10cm, 宽: 32.5cm, 高: 37cm, 净重: 6kg

436 145 玻璃采样瓶

容积 1L, 适用水深 30m, 每个包装 10 只

440 002 塑料包被的使锤

孔径 6mm, 质量 500g

代表文献:

- 1.M. Raateoja¹, J. Seppälä¹, H. Kuosa²,2004.Bio-optical modelling of primary production in the SW Finnish coastal zone, Baltic Sea: fast repetition rate fluorometry in Case 2 waters.Marine Ecology Progress Series.267:9-26.
- 2.Rossberg, Marcelo; Wickham, Stephen A.,2008.Ciliate vertical distribution and diel vertical migration in a eutrophic lake.Fundamental and Applied Limnology / Archiv für Hydrobiologie.171:1-14.
- 3.Hans Ulrik Riisgård, Coralie Barth-Jensen and Caroline V. Madsen,2010.High abundance of the jellyfish *Aurelia aurita* excludes the invasive ctenophore *Mnemiopsis leidyi* to establish in a shallow cove (Kertinge Nor, Denmark).Aquatic Invasions.5(4):doi: 10.3391/ai.2010.5.4.
- 4.David A. Strand, Arne Holst-Jensen, Hildegunn Viljugrein, Bente Edvardsen, Dag Klaveness, Japo Jussila, Trude Vrålstad,2011.Detection and quantification of the crayfish plague agent in natural waters: direct monitoring approach for aquatic environments.Diseases of Aquatic Organisms.95:9-17.
- 5.Iria Durán, Oscar Nieto,2012.Water characterization in three industrialized harbours (Vigo, Bilbao and Pasajes) in North Coast of Spain.Marine Pollution Bulletin.64(2):410-415.
- 6.Christopher O. Miles, Morten Sandvik, Sigrid Haande, Hezron Nonga, and Andreas Ballot,2013.LC-MS Analysis with Thiol Derivatization to Differentiate [Dhb7]- from [Mdha7]-Microcystins: Analysis of Cyanobacterial Blooms, Planktothrix Cultures and European Crayfish from Lake Steinsfjorden, Norway.Environ. Sci. Technol..47(9):4080–4087.